



Homogenization protocol for Strawberry plant leaves using Bullet Blender

Materials:

Samples: 50 mg of fresh leaves

Bead lysis kit: 1.5 mL RINO/Eppendorf Lysis Kit (SKU: NAVYR1/NAVYR5 or NAVYE1/NAVYE5)

Buffer volume: 500 μ L of buffer

Method - Homogenization:

1. Determine the sample size, buffer (depends on the downstream application), buffer volume and the bead lysis kit. **Note:** Choose the correct [lysis kit](#) for optimal homogenization.
2. Cut the sample and place into the buffer-filled tubes (Figure 1).
3. Close the tubes tightly and place into the Bullet Blender. **Note:** Confirm the compatibility of the [contact plate](#) with the tubes (RINO/EPPENDORF) used.
4. Set the speed and time on the Bullet Blender (Table 1). Press "Start", and wait for the run to complete.


Bullet Blender Model	Settings
BB24-AU or BT24M	Speed 12; Time 3 - Dry --> Speed 12; Time 2 - Wet

5. Remove the tubes and visually inspect the samples to confirm complete homogenization (Figure 2).
Note: Foaming in the sample tubes may be observed after homogenization.
6. If homogenization is satisfactory, proceed with the downstream steps.

Figure 1: Pre-homogenization



Figure 2: Post-homogenization



Homogenization verification

